

CLARIFICATIONS AND AMENDMENT

Project Name: **Procurement of Equipment, Goods, Works and Services for Foundation for Innovative New Diagnostics (FIND), India under GFATM financed Project**

Title of Bid : **Limited Tender for Procurement of Biosafety Cabinet- Class II, Type A2 with UPS under RNTCP, Ministry of Health & Family Welfare, Govt. of India**

Amendment No.– 02

Dated :- 27.06.2016

- | | |
|----------------------------|--------------------------------|
| 1. Bid Reference No. :- | SAMS/FIND/BSC/PROCT/LT/08/2016 |
| 2. Date of Notification :- | 13.06.2016 |
| 3. Amendment No. 1 | 20.06.2016 |

Extension of date of submission and opening of Bids (applicable at all relevant sections of tender)

Reference of RFP	Existing	As amended
TERMS & CONDITIONS AND INSTRUCTIONS TO THE BIDDER (ITB)- Clause No. 7 (page nos. 1 and 3)		
Last Date, Time and Place for receipt of Bids	28th June, 2016 till 04.00 PM at Strategic Alliance Management Services Pvt. Ltd. B01-B03, Vardhman Diamond Plaza, Community Centre, D. B. Gupta Road, Paharganj, New Delhi – 110055	4th July, 2016 till 04.00 PM at Strategic Alliance Management Services Pvt. Ltd. B01-B03, Vardhman Diamond Plaza, Community Centre, D. B. Gupta Road, Paharganj, New Delhi – 110055
Last Date, Time and Place for Opening of Bids	28th June, 2016 till 04.30 PM at Strategic Alliance Management Services Pvt. Ltd. B01-B03, Vardhman Diamond Plaza, Community Centre, D. B. Gupta Road, Paharganj, New Delhi - 110055	4th July, 2016 till 04.30 PM at Strategic Alliance Management Services Pvt. Ltd. B01-B03, Vardhman Diamond Plaza, Community Centre, D. B. Gupta Road, Paharganj, New Delhi - 110055

Clarifications to Queries raised by Prospective Bidders

Sl. No.	Reference section of RFP	Queries / Suggestions from Agencies	Clarifications provided by SAMS / FIND
1	Annexure-I (Technical Specification of Bio Safety Cabinet)	You have asked for single piece working surface with integrated (V-) shape front air grill, whereas V shape should not be there.	The requirement in technical specification is as per WHO TB equipment specification and WHO TB Biosafety manual 2012. This technical specification of BSC stands same. Quotation may be submitted accordingly.
2		The required length of the ducting is not mentioned.	Bidder should quote per unit rate for length of ducting in square feet/ square meter with unit rate of required items such as bends, dampers, flanges, DOL starter and other all required materials for proper thimble ducting of BSCs. The unit cost of the quoted ducting materials required for proper installation shall be added in the total cost of the bid and shall be considered for the financial evaluation of the bids
3		Our Equipment is fabricated & tested by NSF. But it is not approved by NSF. Can we quote?	The technical specification of BSC in tender requires only NSF 49 or class II EN 12469 certified BSCs to be quoted

Kindly prepare and submit your bid accordingly.

For Strategic Alliance Management Services Pvt. Ltd.



**Jyoti Singh
Sr. Manager (Procurement)**