

Clarifications regarding queries from prospective Selection of Agencies for the Supply, Installation, Testing and Commissioning of Equipment for TB Laboratories across India under National Competitive Bidding (NCB)

ITB No.: SAMS - FIND- EQUIPMENT PROCT-02/2014

dated 08th February 2014

Sl. No	Query	Clarifications provided by SAMS/ FIND
1	<p>Delivery Period:</p> <p>Delivery period mentioned by you is 45 days. Our standard delivery period is around 12-14 weeks. We therefore request you to amend the delivery period to minimum 12 weeks from the date of firm order received at our end.</p>	<p>The firms were clarified that the issue would be examined and a formal communication would be uploaded</p>
2	<p>Payment Terms:</p> <p>In the tender you have specified that the currency should be Indian Rupees.</p> <p>Please note that the microscopes are produced in abroad and have to be imported from our factory. Hence we would like to offer the price of the product in Foreign currency and request you to place order for direct import by you and to pay in FE directly to our factory. However, we are agreeable to facilitate you with all customs duty and clearance, transport and other related services to move the microscopes to the ultimate destination and shall quote our charges for the same in Indian Rupees, which you can pay to us locally. In case you have any customs duty concession or exemption, please provide us the same and shall charge you accordingly.</p>	<p>The firms were clarified that the issue would be examined and a formal communication would be uploaded</p>
3	<p>Road Permit/ Way Bill etc. for interstate transportation:</p> <p>Please confirm/clarify that the necessary authorization and tax forms/transport permits would be provided by you to enable us to transport the microscopes from the place of import to the ultimate consignee, on behalf of the importer/Buyer.</p>	<p>It was clarified that FIND and SAMS would facilitate the process</p>
4	<p>VAH for Battery Backup of 30 mts</p> <p>Battery VAH has to be mentioned, otherwise the quote from different vendors will be different. Either you give details of Battery AH or VAH Usually for 25.30 mts backup STANDARD UPS WITH need 26 AH Battery as per details:-</p> <p>3Kva UPS with 30 Mts- 26 AH x 8 Battery (VAH = 2496 VAH)</p> <p>5Kva UPS with 30 mts- 26 AH x 16 Battery (VAH = 4680 VAH)</p> <p>For Full 30 mts backup on full load the Battery details are as follows:-</p> <p>3Kva UPS with 30 Mts- 26 AH x 10 Battery (VAH = 3120 VAH)</p> <p>5Kva UPS with 30 mts- 26 AH x 17 Battery (VAH = 5304 VAH)</p>	<p>Bidders may propose the VAH/ no. of batteries as per their design.</p>

<p>5</p>	<p>Battery Warranty of 3 Years The Battery to be used 26 AH does not come with 3 Year Battery Warranty from Manufacturer. Hence it should be 2 Year Only</p>	<p>Bidders may propose the best possible option as per their design.</p>
<p>6</p>	<p>UPS & Battery Warranty from Date of installation UPS and Battery warranty should be wef date of Delivery. Since if the site is not ready and it takes more than a month for installation. There will be problem and the Battery will get discharge and warranty lapses. Hence to a maximum of One month grace from the date of Delivery is expectable.</p>	<p>UPS and Battery warranty would be affected from the date of installation or after one month of delivery whichever is earlier.</p>
<p>7</p>	<p>Battery Make not mentioned in Tender Doc The standard Battery makes should be mentioned like Exide/QUANTA/ROCKET/Hi- Power</p>	<p>Bidders may propose the best possible option as per their design.</p>
<p>8</p>	<p>Some state charges Entry Tax and other Taxes, how will be manage these kind of problem.</p>	<p>It was clarified that FIND and SAMS would facilitate the process. However, any charges if paid would be borne by the vendor.</p>