

**Minutes of Pre-bid Meeting for Procurement of Procurement of Equipment for TB  
Laboratories across India**

**(Bid Ref No.: SAMS/FIND/Proc./Equipment/ATI/13/2019 dt. 26/08/2019)**

The Pre-Bid Meeting was held on 5/09/2019 at 1430 Hrs. in the office of Strategic Alliance Management Services Pvt. Limited (SAMS), New Delhi. Following members were present in the pre-bid meeting:

**SAMS Officials:**

- i. Mr. Satya Verma, GM – Procurement Services.
- ii. Mr. Dinesh Kumar, Procurement Officer
- iii. Mr. Shivam Anand, Manager (Procurement & Hospital Services)
- iv. Ms. Savita Chandna, Procurement Executive

**FIND Officials:**

- i. Dr. DSA Karthickeyan, Medical Officer
- ii. Ms. Madhu Agarwal, Procurement officer
- iii. Mr. Jagdeesh Upreti, Procurement Executive
- iv. Mr. Amit Srivastava, Administrative Assistant – Procurement
- v. Mr. Jagdish Panda, Bio Medical Engineer
- vi. Ms. Sikha Panda, Regional Bio Medical Engineer

**Bidders' representatives:**

Sr. No.	Name and Designation	Name of Firm
1	Mr. Ankit Sakhuja	Ezilt Life Sciences
2	Mr. Ravi Gera, Astd. General Manager-Project	Genaxy Scientific Pvt. Ltd., Delhi
3	Mr. Mukesh	Bio Medicare Laboratories

Proceeding of the pre-bid meeting are as follows:

1. At the outset, SAMS made a briefing about the scope of supplies and related services required through the bid invitation and purpose of the pre-bid meeting.
2. Thereafter, prospective bidders were requested to put up their queries related to scope and terms and conditions given in the Bid Document.
3. The queries from prospective bidders were appropriately responded.
4. The queries sought from prospective bidders as asked during the pre-bid meeting, clarifications sought through mail and purchaser's responses / amendment against each query is summarized in the table given at **Annexure-A**.



**(Dinesh Kumar)  
Procurement Officer**

**Annexure – A**

**Responses / Amendments (Amendment No. 3) with regard to queries/suggestions received for Procurement of Goods & Equipment and Related Services for TB Laboratories across INDIA (Bid Ref No. : SAMS/FIND/Proc./Equipment/ATI/13/2019 dt. 26/08/2019)**

<b>Sl. No.</b>	<b>Para / Clause Reference in Bid Document/ Page No.</b>	<b>Content of Para / Clause under Reference as per Bid Document</b>	<b>Query / Suggestions</b>	<b>Response / Amendments (Amendment No. 3)*</b>
1	3. Technical Specifications; B. Technical Specifications - Specific Requirements <b>Schedule No. 1 Name of Equipment: Precision Balance</b>  Page No. 36-37	2 (a) Main specifications: Electronic balance, weighing capacity 60 g, 120 g or 210 g.	It is requested to amend the required Cap. 210g as mentioning 60g,120g & 210g does not clears what capacity is required.	<b>The requirement is being amended as under:</b>  <b>2a. Main specifications: Electronic balance, weighing capacity 210 g.</b>
	3. Technical Specifications; B. Technical Specifications - Specific Requirements <b>Schedule No. 3 Name of Equipment: Autoclaves: Vertical</b>  Page No. 40-41	2 (b) Main specifications: Triple walled construction; chamber, basket, door lid, doorframe, bolts made of corrosion-resistant material and able to prevent stress cracking preferably made of high grade stainless steel sheet of SS-304 grade, housing with SS legs.  2 (i) Main specifications: Working pressure: 15 PSI, Maximum operating pressure: 2.5 bar or 36 PSI  4 (a) Stainless steel basket (where 2 fit in autoclave directly plus two spare total 4) 4 (b) Stainless steel wire basket (where 2 fit in autoclave directly plus two spare total 4)	Please clarify only Triple Walled required or bidder can quote for Double walled also.  It is requested to amend to Maximum operating pressure 25 PSI to 30 PSI instead of 36 PSI.	2 (b) It is clarified that only Triple walled construction; chamber, basket, door lid, doorframe, bolts made of corrosion-resistant material and able to prevent stress cracking preferably made of high grade stainless steel sheet of SS-304 grade, housing with SS legs is required.  <b>The requirement is being amended as under:</b> <b>2 (i) Main specifications: Working pressure: 15 PSI, Maximum operating pressure: 2.07 bar or 30 PSI</b>

Sl. No.	Para / Clause Reference in Bid Document/ Page No.	Content of Para / Clause under Reference as per Bid Document	Query / Suggestions	Response / Amendments (Amendment No. 3)*
		<b>7 Alarm:</b> audible, with display on dysfunction & after completion of sterilization cycle. Electrical control box, fitted with toggle switch, indicating Neon lamps for Autoclave ON/OFF status, heater ON/Off status. Over-temperature and over-pressure protection limiter	display on dysfunction creates confusion so it is requested to kindly removed from the same from Technical specification.	No change in technical specifications.

*\*the bold contents are the amendments (**Amendment no. 3**) against the respective paras of Bid Document. **The referred amendments shall be applicable to all relevant Sections of the Bid.***